


PROPRIETARY INFORMATION:
 PROPRIETARY OF GLOBTEK, INC. ANY REPRODUCTION, DISCLOSURE OR USE OF THIS DRAWING, IN WHOLE OR IN PART, IS HEREBY PROHIBITED EXCEPT AS SPECIFIED IN WRITING BY GLOBTEK, INC.

REVISION			
REV	DESCRIPTION	DATE	APPROVED
A	INITIAL RELEASE	04/09/01	PS
B	REDRAWN, , RFS.2385, FS	10/14/03	BS
C	REVISE APPROVALS , RFS.3744, DMR	05/25/04	BS




GENERAL SPECIFICATIONS:

RATING: 13A / 125V AC
 CABLE TYPE / SIZE: SJT, 16 AWG X 3 CONDUCTORS
 TEMPERATURE RATING: 60° C MIN.
 CONDUCTOR COLORS: BLACK / WHITE / GREEN
 CABLE JACKET COLOR: GRAY (GA001 SIMILAR TO PMS 423C)
 OVERALL LENGTH (L): 3050 ± 25 mm (10 Ft)
 CABLE OUTER DIAMETER: 8.5 ± 0.15 mm
 END 'A' TERMINATION: HOSPITAL GRADE NEMA 5-15P CLEAR
 END 'B' TERMINATION: IEC 320 / C13 CONNECTOR, CLEAR
 APPROVALS PLUG: UL, CSA or cUL
 APPROVALS CABLE: UL, CSA or cUL
 APPROVALS CONNECTOR: UL, CSA or cUL

PACKING:

HANKING
 L1 = 0
 L2 = 210 ± 10 mm
 L3 = 0



NOTES:

1. CORD, PLUG AND CONNECTOR MUST BE MARKED WITH APPROVAL SYMBOLS OR FILE NUMBER(S).



2. BOTH PLUG AND CONNECTOR ARE CLEAR IN COLOR.
 3. THE CORD JACKET IS GRAY IN COLOR.

DASH NO.	PART NO	REV.	DESCRIPTION	NOTES
TABULATION BLOCK				
REV.	C			
SHEET	1			
INIT. BY: FS	DATE: 10/14/03	TOLERANCES: DECIMALS ANGULAR +/- 1° .xx +/- 0.1 .xxx +/- .005 MILLIMETERS ANGULAR +/- 1° .xx +/- 0.3 .xxx +/- .013		
DRAWN: NF	DATE: 10/14/03	FSCN No.: G, CC, YL	SIZE A	GlobTek, Inc. 186 Veterans Dr. Northvale, NJ 07647 Tel. 201-784-1000 Fax 201.784.0111 www.globtek.com DWG TITLE: HOSPITAL GRADE NORTH AMERICAN DETACHABLE CORDSET
APRVD: DMR	DATE: 10/14/03	SCALE: NONE	PART NO: 11910510F701/GY	
Foot Note: Globtek Inc. will not be liable for the safety and performance of these power product if unauthorized access and repair occurs. End user should consult the applicable UL, CSA or EN Standards for proper installation instruction.		Limitation of Use: Globtek product are not authorized for use as mission critical components in life support, hazardous environment, nuclear or aircraft applications without prior written approval from the CEO of Globtek Inc.		
Contents of this document are subject to change without prior notice		REV. C SHEET 1 / 1		